

EC Declaration of Product Conformity

The manufacturer established in the community;

BM Polyco Ltd
Crown Road
Enfield
EN1 1TX
United Kingdom

Certified to ISO 9001 & ISO 13485

declares that the new medical device described hereafter



BLACK NITRILE GL897

Non-sterile, powder-free, nitrile disposable glove

is in conformity with the provisions of Council Directive 93/42/EEC and with National Standards transposing harmonised standards EN455-1, EN455-2, EN455-3, EN455-4 and is self-certified as a Class 1 non-sterile medical device

Freedom from holes: AQL 1.5
Physical properties: 6N minimum

is in conformity with the provisions of Council Directive 89/686/EEC and where such is the case, with National Standards transposing harmonised standards EN374-1:2003, EN374-2:2003, EN374-3:2003 and EN420:2003.

is identical to the PPE which is the subject of EC certificate of conformity no. 1572 issued by SATRA Technology Centre, Wyndham Way, Telford Way, Kettering, Northamptonshire, NN16 8SD, United Kingdom (Notified Body No. 0321).

conforms to the provisions of Council Directive 89/686/EEC Article 11A under the supervision of SATRA (Notified Body No. 0321).

**MICRO ORGANISM
HAZARDS
EN374-2**



**CHEMICAL
HAZARDS
EN374-3**



CE 0321

Complex Design



User Information

This product does not contain natural rubber latex.
Glove Care: Store below 40 °C away from direct sunlight.
You are advised to retain this packaging for reference.
Contains accelerators which may cause allergic reactions.

Range Available: sizes and codes

Size	Small	Medium	Large	X-Large
Code	GL8971	GL8972	GL8973	GL8975

Food Contact



GL897 have been tested in accordance with EN1186. They are suitable for use with all food categories in situations of short term, repeat contact.

Done at Enfield, 07/08/14

Bernard Garvey
Technology Director