

EC Declaration of Product Conformity

The manufacturer established in the community;

BM Polyco Ltd
Crown Road
Enfield
EN1 1TX
United Kingdom

Certified to ISO 9001 & ISO 13485

declares that the new Medical Device and PPE described hereafter



FINITE P INDIGO AF

Non-sterile, powder-free, synthetic nitrile rubber
disposable examination glove

is in conformity with the provisions of Council Directive 93/42/EEC and with National Standards transposing harmonised standards EN455-1, EN455-2, EN455-3, EN455-4 and is self-certified as a Class 1 non-sterile medical device

Freedom from holes: AQL 1.5
Physical properties: 6N minimum

is in conformity with the provisions of Council Directive 89/686/EEC and where such is the case, with National Standards transposing harmonised standards EN374-1:2003, EN374-2:2003, EN374-3:2003 and EN420:2003.

is identical to the PPE which is the subject of EC certificate of conformity No. 3766 issued by SATRA Technology Centre, Wyndham Way, Telford Way, Kettering, Northamptonshire, NN16 8SD, United Kingdom (Notified Body No. 0321).

conforms to the provisions of Council Directive 89/686/EEC Article 11B under the supervision of the Notified Body, LRQA Ltd, Hiramford, Middlemarch Office Village, Siskin Drive, Coventry, CV3 4FJ (Notified Body No. 0088).

**MICRO ORGANISM
HAZARDS
EN374-2**



**CHEMICAL
HAZARDS
EN374-3**



CE 0088



Complex Design

User Information

This product does not contain natural rubber latex.
This product does not contain chemical accelerators.

Glove Care

Storage: Store below 40°C in dry conditions away from direct sunlight and heat.

You are advised to retain this packaging for reference

Range Available: sizes and codes

Sizes	5.5	6.5	7.5	8.5	9.5
Code	MFNP100/00	MFNP100/01	MFNP100/02	MFNP100/03	MFNP100/04

Food Contact



Finite P Indigo AF have been tested in accordance with EN1186. They are suitable for use with all food categories in situations of short term, repeat contact.

Done at Enfield, 27/01/14

Bernard Garvey
Technology Director